Quality Certificate

VWR® ERLENMEYER FLASKS

As per the manufacturer, the below product meets the following criteria:

North American Catalog No	89095-284
Lot Number	02518001
Description	VWR FLASK PC 1L STERILE CS25

Component Materials:

Flask	Virgin Polycarbonate, meets USP, Class VI requirements for plastic containers and closures.
Cap	Virgin Polypropylene, meets USP, Class VI requirements for plastic containers and closures. Heavy metal free (meets CONEG requirements) color concentrate.

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:

Test/Procedure	**Results**
Leak Test	Pass
Visual Attributes	Pass
Packaging	Pass

Sterilization: The lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 (TIR 33) "*Sterilization of health care products-Requirements for validation and routine control-Radiation sterilization*". **Sterility Assurance Level:** SAL 10^{-6}

RNase/DNase Testing: This product has been tested and is free of any detectable RNase/DNase contamination.

Pyrogens: The product has been tested and has met the criteria established in ANSI/AAMI ST 72:2002/(R) 2010: "*Bacterial Endotoxins - Test methodologies, routine testing, and alternative to batch testing".*

Lot Number Designation: 8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Run number for that year.

Signed:

Linda K. Shine
Manager - Supplier Qualification
Global Compliance
VWR International LLC

Date: February 21, 2018