



Quality Certificate

VWR® CLINICAL 50 CENTRIFUGE

This document certifies that the below product meets the following criteria:

North American Catalog No:	82013-800
Description:	VWR CENTRIFUGE CLINICAL 50 120V

- **QC testing** process &/or Industry standards : 100% Quality Control process verifies speed and timer accuracy.
- **Accuracy** - (+/- 500 rpm)
- **Country of origin** as defined by International standards - China

Signed:

A handwritten signature in blue ink, appearing to read "Ken Crossley", with a stylized flourish at the end.

Ken Crossley
Manager
Quality Assurance

Date: March 19, 2015